




Doc. No. 3023
Date 31/07/2026

Version 5.0
Page 1 of 1
Document Owner PROD-BGS

Title **Cellbind P3 release**

K7210



Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 27 August 2026 Approved by QA



CONFIDENTIAL

Use copy according to Doc. ID4016, Printed on: 17/08/2026