

Instructions for use



Essange Reagents
Plesmanlaan 125
1066 CX Amsterdam
The Netherlands

Phone: +31 20 4570707

info@essange-reagents.nl
www.essange-reagents.com

PeliKine Human IL-6 ELISA Rapid Set A

REF M2018

336 version **07** Issued **21-Sep-2026** (en)

RUO

**RUO-For Research Use
Only. Not for use in
diagnostic procedures.**

General Information

The PeliKine Human IL-6 ELISA Rapid Set is an enzyme linked immunoassay (ELISA) for fast, reproducible and specific quantitative determination of IL-6 in MAT supernatant. This ELISA was validated in combination with Essange Reagents' monocyte activation test (MAT) kits: MAT Cell Set (M2016) and MAT Cell Set HS (M2017). The MAT is an in vitro-based assay which determines pyrogenic activity in pharmaceutical products by measuring cytokine production of human monocytes. Pharmaceutical products, intended for parenteral use, must be free of pyrogens (fever-inducing substances), which can originate from bacteria, viruses and fungi. The MAT detects endotoxin and non-endotoxin pyrogens that are relevant to humans. The European Pharmacopoeia (Ph. Eur. chapter 2.6.30) is the main guideline regarding MAT procedure and product-specific validation. Note that the PeliKine Human IL-6 ELISA Rapid Set consists of 2 kit parts: PeliKine Human IL-6 ELISA Rapid Set A (M2018) and PeliKine Human IL-6 ELISA Rapid Set B (M2019). The contents of this IFU (except for sections "Package contents" and "Precautions") are identical to the contents of the IFU for the PeliKine Human IL-6 ELISA Rapid Set B (M2019)s

Package Contents

- Leaflet 336-version_07

Name	Amount	Concentration	Appearance	Ref
IL-6 standard	1 x 750 µL	See label for concentration	Black cap	M191603
Biotinylated antibody	1 x 375 µL	100-fold concentrated	Yellow cap	M191604
Streptavidin-poly-HRP conjugate	1 x 20 µL	10,000-fold concentrated	Brown cap	M1942
HPE-dilution buffer	1 x 55 mL	5-fold concentrated	White bottle	M1940

- Consult the Certificate of Analysis for additional information.

Materials Required (not provided)

The following equipment and materials are required but not provided:

- Microplate photometer for measuring absorbance at 450 nm, with a reference wavelength set between 540 and 590 nm
- Reagents reservoirs
- Distilled or deionised water
- Calibrated pipettes (5-1000 µL)
- Calibrated multichannel pipettes (30-300 µL)
- Beakers, flasks, cylinders, tubes and liquid containers necessary for preparation of reagents
- PeliKine Human IL-6 ELISA Rapid Set B (M2019)

Optional

- Automated microplate washer
- U bottom plate

Specimen Collection and Handling

Cell culture supernatant should be collected and handled according to standard cell culture procedures, see also IFU MAT Cell Set (M2016) and MAT Cell Set HS (M2017).

Instructions for use



Test Principle

The PeliKine Human IL-6 ELISA Rapid Set is a "sandwich-type" of enzyme immunoassay. This kit includes the proven antibodies (capturing [i.e. coating antibody] and biotinylated antibody) as found in the PeliKine Human IL-6 ELISA kit (M1916)². Samples, containing IL-6 (i.e. MAT supernatant), together with a biotinylated anti-human IL-6 antibody, are added to the microtiter plate. The microtiter plates are pre-coated with anti-human IL-6 capturing antibody. In the microtiter plate, IL-6 is simultaneously captured to polystyrene microtiter wells and bound by the biotinylated anti-human IL-6 antibody. Non-bound material is then removed by washing. Subsequently, streptavidin-horseradish peroxidase (HRP)-conjugate is added, which binds to the biotinylated antibody. After removal of non-bound HRP conjugate by washing, substrate solution is added to the wells. A coloured product is formed in proportion to the amount of IL-6 present in the supernatant. After the reaction has been terminated by the addition of a stop solution, absorbance is measured in a microtiter plate reader.

Calibrators and Controls

The IL-6 standard (M191603) is calibrated to the 2nd International Standard for Interleukin-6 21/308 from the National Institute for Biological Standards and Control (NIBSC).

Storage and Shelf-life

Product PeliKine Human IL-6 ELISA Rapid Set A (M2018) should be stored directly upon arrival, and after use, at -32 to -18°C in the original packaging (preferably shielded from light). Product should be stored upright. When stored under the recommended conditions, the shelf-life, mentioned on the label, is guaranteed.

Precautions and Warnings

- For Research Use Only, not for use in diagnostic procedures. The PeliKine Human IL-6 ELISA Rapid Set is not intended for the detection of IL-6 in clinical samples or as an aid in the diagnosis of human disease.
- The PeliKine Human IL-6 ELISA Rapid Set A (M2018) should be stored at -32 to -18°C.
- All kit components must be stored in the original packaging to allow for verification of the expiration date.
- Reagents (unopened or opened) cannot be used beyond the expiration date, which is printed on the label of each component.
- Leaking or damaged vials or bottles or other materials cannot be used.
- Only use the reagents supplied with the set, do not mix reagents from different kit lots.
- Reagents or remnants of reagents (e.g. dead volume) cannot be mixed with contents of freshly opened vials.
- Caps and vials/bottles are not interchangeable, caps must be replaced on the corresponding vials/bottles.
- Do not use reagents with any evidence of turbidity or microbial contamination.
- Reagents cannot be assumed to be free from infectious agents.
- Handle all samples with care to prevent transmission of blood-borne infections.
- Sodium azide inactivates HRP, do not use sodium azide-containing solutions, nor add sodium azide to the supplied materials.
- IL-6 standard and biotinylated antibody contain 0.001% Merthiolate, and HPE-dilution buffer contains 0.005% Merthiolate. Do not add any preservative to the supplied reagents, they may have direct or indirect effects on the final colour development of the HRP-substrate system.
- Salt crystals may appear in the 5x concentrated HPE buffer. Before preparing the working-strength buffer, first warm the concentrated buffer BRIEFLY to 37°C in a water bath to dissolve the precipitate.
- HPE buffer contains EDTA, refer to the MSDS of M2018 for details.
- Avoid repeated freeze-thawing of the IL-6 standard, although experimental data have shown that up to 3 freeze-thaw cycles have no effect on the IL-6 levels of the standard.
- Prior to the assay, frozen supernatant/samples should be thawed as quickly as possible in tap water (18-25°C), do not use 37°C or 56°C water baths for this purpose.
- Care must be taken in the use and disposal of each container and its contents. Waste disposal, after completion of the test, must be performed according to laboratory regulations.
- For optimal performance of ELISA be sure that all pipets and systems are checked and under full maintenance service according to described procedures of the manufacturers.

Material safety data sheet

The Material safety data sheet for this product can be found on our website: www.essange-reagents.com/ by searching for the product number.

Instructions for use



Abbreviations

ELISA	Enzyme-Linked Immunosorbent Assay
HPE	High Performance ELISA
HRP	Horseradish peroxidase
HS	Human serum
IL-6	Interleukin-6
LOD	Limit Of Detection
MAT	Monocyte Activation Test
Ph. Eur.	European Pharmacopoeia
TMB	3,3',5,5'-Tetramethylbenzidine

Assay Procedure

Components marked with “^(a)” are part of the PeliKine Human IL-6 ELISA Rapid Set A (M2018), components marked with “^(b)” are part of the PeliKine Human IL-6 ELISA Rapid Set B (M2019).

General guidelines

1. Allow all reagents to calibrate to room temperature (18–25°C) prior to use, with the exception of the streptavidin-HRP conjugate which has to be kept at –32°C to –18°C to ensure stability.
2. The complete assay must be performed at room temperature (18–25°C) without shaking. Shaking is optional, step 5 (incubation of supernatant and biotinylated antibody) and step 8 (incubation of streptavidin-HRP conjugate) of section ‘Procedure’ can be performed on a horizontal plate shaker (700 ± 100 rpm).
3. The stated volumes are for *one* ELISA plate.
4. If more than one plate is assayed, first complete the steps of adding supernatant and biotinylated antibody (as described in the second and third points of section “procedure” step 4) for the first plate, then repeat and complete these steps for each plate.
5. Mix all reagents thoroughly but gently before use (without foaming).
6. Centrifuge all vials before use (1 minute 3000 x g).
7. Do not allow wells to stand uncovered or dry for extended periods between incubation steps.
8. Carefully remove all air bubbles from the wells before incubation.
9. The opalescent working-strength HPE-dilution buffer can be stored for up to one week at 2–8°C.
10. The working-strength wash buffer can be stored up to 2 months at 2–8°C.
11. Protect TMB substrate solution from prolonged exposure to light.
12. The procedure in section “procedure” is designed for a 5x supernatant dilution, for examples of preparation of other dilutions refer to section “Examples of supernatant dilutions”.

Procedure

1. Prepare working-strength HPE-dilution buffer:
 - Add 15 mL 5-fold concentrated HPE-dilution buffer^(a) to 60 mL distilled water.
2. Prepare IL-6 standard in working-strength HPE-dilution buffer: (Optional: An IL-6 standard curve may be included in the ELISA as a quality control, but is not necessary to determine the pyrogenic activity of the sample.)
 - Label 7 tubes, one tube for each dilution: 2250, 750, 250, 83.3, 27.8, 9.3, 3.1 pg/mL (final concentrations in the ELISA plate are: 450, 150, 50, 16.7, 5.6, 1.9, 0.6 pg/mL)
 - Prepare the highest dilution by pipetting IL-6 and working-strength dilution buffer into the tube labelled 2250 pg/mL, minimum 100 µL total volume, mix well.
(Example: If stock concentration is 4000 pg/mL, then 63 µL IL-6 standard is mixed with 49 µL HPE buffer)
 - Add 80 µL of working-strength dilution buffer into the other tubes.
 - Transfer 40 µL of the highest dilution into the second tube labelled 750 pg/mL.
 - Repeat the serial dilutions five more times by adding 40 µL of the previous tube of diluted standard to the next tube containing 80 µL of dilution buffer.
 - It is recommended to prepare two separate series (duplicate) for each assay.
3. Prepare the working-strength biotinylated antibody (81x diluted, final dilution in the ELISA plate is 101x):
 - Add 120 µL biotinylated antibody^(a) to 9.6 mL HPE buffer. Mix by inverting the tube 10 times.
 - Add this mix to a reagents reservoir.
4. Add the supernatant and the working-strength biotinylated antibody to the IL-6 ELISA pre-coated plate:
 - Homogenise the harvested supernatant by pipetting up and down for 5 times and add 20 µL to the corresponding wells in the IL-6 ELISA pre-coated plate.

Instructions for use

- Add 20 µL of the prepared IL-6 standard to columns 11 and 12 of the pre-coated plate and 20 µL working-strength HPE-dilution buffer to the blank wells (see also Figure 1).
 - Immediately add 80 µL working-strength biotinylated antibody to the wells, containing supernatant and standard curve, with a multichannel pipette and mix 3x gently by pipetting up and down; change the pipetting tips each row.
 - Cover the pre-coated plate with a seal and **incubate for 1 hour at 18-25°C**.
5. Prepare wash buffer:
 - Prepare working-strength wash buffer by adding 50 mL of the wash buffer concentrate^(b) (total content of the bottle) to 950 mL distilled water.
 6. Just before the washing step, prepare the working-strength streptavidin-HRP conjugate:
 - Add 2 µL 10,000-fold concentrated streptavidin-poly-HRP conjugate^(a) to 20 mL working-strength HPE-dilution buffer. Mix by inverting the tube 10 times.
 7. After 1 hour of incubation wash the plate 5 times with wash buffer (volume > 300 µL) manually or on a plate washer. After the final emptying the plate should be dry.
 8. Add 100 µL of working-strength streptavidin-HRP conjugate to the appropriate wells and cover the microplate with adhesive seal. Gently agitate the microplate by tapping the edge of the plate for a few seconds to mix contents of each well and **incubate for 30 minutes at 18-25°C**.
 9. After 30 minutes of incubation wash the plate 5 times with wash buffer (volume > 300 µL) manually or on a plate washer. After the final emptying the plate should be dry.
 10. Add 100 µL TMB substrate solution^(b) to all wells. Gently agitate the microplate by tapping the edge of the plate for a few seconds to mix contents of each well and **incubate for 10 minutes (recommended) at 18-25°C in the dark** (do not use aluminium foil, but place the microplate e.g. in a cupboard).
 11. After 10 minutes of incubation add 100 µL of stop solution^(b) to all wells. **After stopping the substrate reaction the colour is stable for maximally 30 minutes.**
 12. Place the microplate in an ELISA reader and record absorbance at 450 nm, with a reference wavelength set between 540 and 590 nm (subtract the reference wavelength measurement from the measurement at 450 nm if the software does not subtract this automatically).

	1	2	3	4	5	6	7	8	9	10	11	12
A	Cell culture supernatant										450 pg/mL	
B											150 pg/mL	
C											50 pg/mL	
D											16.7 pg/mL	
E											5.6 pg/mL	
F											1.9 pg/mL	
G											0.6 pg/mL	
H											Blank	

Figure 1: Suggested plate lay-out for the ELISA. Depicted concentrations of the IL-6 standard curve are final concentrations in the ELISA plate.

Examples of supernatant dilutions

If a supernatant dilution other than 5x is preferred, then pre-dilute the supernatant (in e.g. a U bottom plate) according to the examples in the table below. Mix the supernatant, together with the working-strength HPE-dilution buffer, 5x and then transfer 20 µL pre-diluted supernatant to the IL-6 ELISA pre-coated plate.

Dilution	Supernatant	1x HPE buffer
10x	30 µL	30 µL
20x		90 µL
50x		270 µL

Results

Consult Ph. Eur. chapter 2.6.30 and the IFUs for MAT Cell Set (M2016) and MAT Cell Set HS (M2017) for guidelines on data analysis.

Interpretation

Consult Ph. Eur. chapter 2.6.30 and the IFUs for MAT Cell Set (M2016) and MAT Cell Set HS (M2017) for guidelines interpretation of results.

Instructions for use



Specifications

Consult the Certificate of Analysis on performance characteristics.

Limitations

- This ELISA kit has been designed for professional use only and for analysis of supernatant obtained from the MAT Cell Set (M2016) and MAT Cell Set HS (M2017). The user must be trained and familiar with pyrogen test and ELISA test procedures.
- Only the washing steps were validated by Essange Reagents on an automated microplate washer (Biotek 405 TS). The other steps, as described in this IFU, were validated by Essange Reagents with a manual testing procedure. All claims in this IFU were validated with the manual testing procedure (except for the washing steps). When using the kit on an ELISA automate, the test must be validated by the user before use. The claims in this IFU are not valid for the performance of this kit on an ELISA machine.
- The ELISA procedure was validated for a 5x MAT supernatant dilution. Other dilutions must be validated by the user.
- This kit was validated with supernatant obtained from the MAT Cell Set (M2016) and MAT Cell Set HS (M2017). Unlike M1916, this kit was not validated for analysis of IL-6 in serum and plasma samples. The use of supernatant from other cell or supplement sources or other matrices must be validated by the user.

References

1. European Directorate for the Quality of Medicines. *Ph. Eur.* chapter 2.6.30 (07/2024): Monocyte Activation Test.
2. Helle M, Boeije L, de Groot E, de Vos A, Aarden L; Sensitive ELISA for interleukin-6. Detection of IL-6 in biological fluids: synovial fluids and sera. *J.Immunol.Methods*. 1991 Apr 8;138(1):47-56.

Disclaimer

FOR RESEARCH USE ONLY,

Essange Reagents products are guaranteed to perform as described in the original manufacturer's instructions for use. Essange Reagents declines all responsibility arising from any deviation thereof. The user is responsible for proper calculations and subsequent interpretations of the results. Essange Reagents is not liable for any claims, including third party claims, due to or caused by pyrogenic activity in pharmaceutical products. Consult the relevant chapters in the European Pharmacopoeia (Ph. Eur. 2.6.30) as well as other documents relevant for your specific purpose for utilization of the test.

Changes to Previous Version

The IFU has been updated to align with the current corporate design.

The dilution procedure has been revised to accommodate batch-to-batch variation in the IL-6 standard concentration. The concentration is now specified on the product label, and no fixed dilution volume is provided.