
K1189

Clone number: LB2



Analysis	Requirement			Result	
Specificity	Negative with cells negative for the B antigen			Conform requirements	
Avidity		B	A_1B		
	Tube method	≥ 3+	≥ 3+	Conform requirements	
	Microplate method	≥ 3+	≥ 3+	Conform requirements	
	Cellbind Screen method	≥ 3+	≥ 3+	Conform requirements	
	Cellbind Direct Type method	≥ 3+	≥ 3+	Conform requirements	
Titer		B	A_1B		
	Tube method	≥ 1:128	≥ 1:32	Conform requirements	
	Microplate method	≥ 1:128	≥ 1:32	Conform requirements	
Growthtest	no growth			Conform	
External	released by Notified Body			Released	(Y) / ~~N~~

Date : 17 November 2026 Approved by QA/QC :