
Sanquin Reagents	Doc. Nr. REA-S-0021	Revisie	03
	Datum 13 Feb 2025	Pagina	1 van 1
		Document Eigenaar	PROD-QC

Titel **Pelicontrol release**

K1379



Analysis	Requirement	Result
Control cells	DAT* negative	Conform requirements
	sample 1:	
	A_2B Rh D positive, negative screening.	Conform requirements
	sample 2:	
	O Rh D negative, positive screening	Conform requirements
Hematocrit	20 ± 2%	Conform requirements

* = direct antiglobulin test

All the donor units and sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 12 November 2026 Approved by QA/QC :