


Title **MASPAT Indicator Red Cells release**



Analysis	Requirement				Result
Antigen control		anti-A	anti-B	DAT	
	MIRC 0.3%	0	0	≥ 2+	Conform requirements
Hematocrit	MIRC 0.3 ± 0.1%				Conform requirements

The donor unit used in the production of this reagent has been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Approved by

Date: 20 January 2026

QA: 