K1190
Clone numbers: ES4/ES15



Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
VL140/c
Pelikloon
anti-A,B (IgM) monoclonal
Clones: ES4, ES15
10mL
IVD 2-8°C REF K1190 CE 0344
LOT **8100100182** **2028-10-06**

Analysis	Requirement				Result
Specificity	Negative with O cells				Conform requirements
Avidity		A_1	A_2	B	
	Tube method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
	Microplate method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
	Cellbind Screen method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
	Cellbind Direct Type method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
Titer		A_1	A_2	B	
	Tube method	≥ 1:256	≥ 1:64	≥ 1:128	Conform requirements
	Microplate method	≥ 1:256	≥ 1:64	≥ 1:128	Conform requirements
Growthtest	no growth				Conform
External release	released by Notified Body				Released Y / ~~N~~

Date : 17 November 2026 Approved by QA/QC :