

Titel **MASPAT kit release**

K1360

 Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

MASPAT kit

Monoclonal Antibody Solid Phase

Platelet Antibody Test


IVD Σ 96

LOT 8000462145

REF K1360



2-8 °C

2027-06-15

(01)1872068203
0025(10)800046
2145(17)270615

Component	Requirement	Result
Positive control Lot: <u>8000461978</u>	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements Conform requirements
Negative control Lot: <u>8000461920</u>	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: <u>8000461954</u>	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: <u>8000461923</u>	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: <u>8000461871</u>	Conductivity 3.4 – 4.8 mScm ⁻¹ pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 22 January 2016 Approved by QA/QC : 

VERTROUWELIJK

Use copy according to REA-SOP-0002

Printed on 12 Jan 2026 at. 07 17 44 am , Printed by M DEMOL