






Titel **MASPAT kit release**





K1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

MASPAT kit
Monoclonal Antibody Solid Phase
 Platelet Antibody Test



  96
 **LOT** 8000462355
 **REF** K1360

   2-8 °C
 2027-09-30 (01)1872068203
 0025(10)800046
 2355(17)270930

Component	Requirement	Result
Positive control Lot: 8000461978	Positive reaction with: platelets Titration with: platelets \geq 1:4	Conform requirements Conform requirements
Negative control Lot: 8000461979	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000461954	Titration with: platelets \geq 1:4	Conform requirements
Microplates Lot: 8000462093	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000462258	Conductivity 3.4 – 4.8 mScm ⁻¹ pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 23 APRIL 2026 Approved by QA/QC : 