K1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL
CL

MASPAT kit
Monoclonal Antibody Solid Phase
Platelet Antibody Test

www.sanquin.org/reagents

IVD Σ 96
LOT 8100100249
REF K1360

2-8 °C
2028-01-26

(01)1872068203
0025(10)810010
0249(17)280126

Component	Requirement	Result
Positive control **Lot:** 8000462417	Positive reaction with: platelets Titration with: platelets ≥ 1:4	 Conform requirements Conform requirements
Negative control **Lot:** 8000462418	Negative reaction with: platelets	 Conform requirements
Anti-IgG **Lot:** 8100100076	Titration with: platelets ≥ 1:4	 Conform requirements
Microplates **Lot:** 8100100153	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	 Conform requirements Conform requirements Conform requirements
LISS **Lot:** 8100100077	Conductivity 3.4 – 4.8 $mScm^{-1}$ pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 12 November 2026 Approved by QA/QC :