

Titel **Pelicontrol release**

K1379



Analysis	Requirement	Result
Control cells	DAT* negative  <b>sample 1:</b>  A <sub>2</sub> B Rh D positive, negative screening.	Conform requirements  Conform requirements
	<b>sample 2:</b>  O Rh D negative, positive screening	Conform requirements
Hematocrit	20 ± 2%	Conform requirements

\* = direct antiglobulin test

All the donor units and sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date

20 January 2026 Approved by QA/QC

