

Titel **Makropanel 16 (P) release**

K1385

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

Makropanel 16
3 ± 1%
16 x 3 mL

IVD
LOT 8000462256
REF K1385

www.sanquin.org/reagents
CE 0344 UDI
2-8 °C
2026-04-27

(01)1872068203
1077(10)800046
2256(17)260427



Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	3 ± 1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 13 March 2026 Approved by QA/QC

