



Doc. Nr. REA-S-0010
Datum 23 Jun 2025

Revisie 03
Pagina 1 van 1
Document Eigenaar PROD-QC

Titel **cellbind P2 release**

K7200



Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 12 march 2026 Approved by QA/QC : 