



Doc. No. 3017
Date 06/07/2026

Version 4.0
Page 1 of 2
Document Owner PROD-BGS

Title cellbind ID 16 (P) release

K7230

 **Sanquin Plesmanlaan 125 1066 CX Amsterdam NL**

Cellbind ID16
0.5 ± 0.1 %

16 x 3 mL ^{CL}



IVD
LOT 8100100109
REF K7230

 www.sanquin.org/reagents

CE **UDI** 0344

 2-8 °C
2026-10-12

(01)1872068203
1107(10)810010
0109(17)261012

Analysis	Requirement	Result
Antigen control	Various antigens have been tested and found conform the enclosed antigram of the panel	Conform requirements
Hematocrit	0.5 ± 0.1%	Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

Date : 28 August 2026 Approved by QA



CONFIDENTIAL

Use copy according to Doc. ID4016, Printed on: 17/08/2026