




Doc. No. REA-S-0022
Date 30 May 2025

Revision 04
Page 1 of 1
Document Owner PROD-QC

Title Cellbind reagent red cells release



Analysis	Requirement					Result
Antigen control		anti-A	anti-B	anti-D	DAT	
	A ₁ *	≥ 3+	0	n.t.**	0	Conform requirements
	B	0	≥ 3+	n.t.**	0	Conform requirements
Hematocrit	0.5 ± 0.1%					Conform requirements

All the donor units used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with current EU guidelines to GMP and found to be non reactive.

* Blood group 'A subtype' has been established in another technique and the results are conform specification.

** n.t.: not tested.

Approved by

Date: 03 September 2026

QA:

CONFIDENTIAL

Use copy according to Doc. ID4016, Printed on: 24/08/2026