

Title **Certificate of Analysis PeliKine Human IL-6 ELISA Rapid Set A (M2018)**

Product information

Product name	PeliKine Human IL-6 ELISA Rapid Set A		
Product number	M2018		
Batch number	8000462279		
Label	<div style="display: flex; justify-content: space-between;"> <div style="width: 60%;"> <p style="font-size: small; background-color: #333; color: white; padding: 2px;">Sanquin Reagents B.V. Plesmanlaan 125 1066 CX Amsterdam NL</p> <h2 style="margin: 0;">PeliKine Human IL-6 ELISA Rapid Set A</h2> <p style="margin: 5px 0;">Research Use Only Spin small volumes before use</p> <p style="margin: 5px 0;"> -18 to -32°C LOT 8000462279 2028-04-10 REF M2018 </p> </div> <div style="width: 35%;"> <p>USA Warning Causes serious eye irritation May cause damage to organs through prolonged or repeated exposure Contains tetrasodium ethylenediaminetetraacetate Do not breathe dust/fume/gas/mist/vapors/spray Wear eye protection/face protection If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Get medical advice/attention if you feel unwell. If eye irritation persists. Get medical advice/attention. Dispose of contents/container in accordance with local/regional/national/international regulations.</p> <p>EU Warning Causes serious eye irritation. Wear protective gloves/protective clothing/eye protection/face protection. IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. If eye irritation persists: Get medical advice/attention. Tel: +31 20 512 3599 Mon-Fri 9:00-17:00</p> </div> </div>		

Test results

Description	Specification	Meets specification
Production and release of product	According to Essange Reagents' certificated ISO13485 quality management system	yes
ELISA	IL-6 standard curve: <ul style="list-style-type: none"> - Tested concentrations: blank, 0.6, 1.9, 5.6, 16.7, 50, 150 and 450 pg/mL - Sensitivity (based on mean of blank + 3x standard deviation): ≤ 1 pg/mL - Mean OD450-550nm of the blank: ≤ 0.140 	yes

Date: 16 March 2026

Approved by QA: