Doc No.	3972	Version	3.0
Date	22/09/2026	Page	1 of 1
		Document Owner	PROD-P&I

essange reagents

Title: Certificate of Analysis PeliKine Human IL-6 ELISA Rapid Set B (M2019)

Product information

Product name	PeliKine Human IL-6 ELISA Rapid Set B
Product number	M2019
Batch number	8100100341

Label

Essange Reagents Plesmanlaan 125 1066 CX Amsterdam NL

CL

PeliKine Human IL-6 ELISA Rapid Set B

RUO For Research Use Only, not for use in diagnostic procedures

LOT 8100100341

REF M2019

2-8 °C

2028-10-21

USA **Danger.** Causes severe skin burns and eye damage. May cause cancer.
Contains: sulfuric acid
Do not handle until all safety precautions have been read and understood. Do not breathe dusts or mists. Wear eye protection/face protection. If swallowed: Rinse mouth. Do NOT induce vomiting. If on skin (or hair): Take off immediately all contaminated clothing. Rinse skin with water/shower. If inhaled: Remove person to fresh air and keep comfortable for breathing. If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a poison center/doctor. Specific treatment (see on this label). Wash contaminated clothing before reuse Store locked up. Dispose of contents/container in accordance with local/regional/national/international regulations.

EU Safety data sheet available on request.
Tel: +31 20 512 3599 Mon-Fri 9:00-17:00 CET

Test results

This product is tested according to Essange Reagents release procedures

Description	Specification	Meets specification
ELISA	IL-6 standard curve • Tested concentrations 0, 0.6, 1.9, 5.6, 16.7, 50, 150, 450 pg/mL • Mean $OD_{450-550nm}$ of the blank < 0.1 • Limit of Detection (mean blank + 3x SD) ≤ 1 pg/mL Plate homogeneity CV% ≤ 7.5	Yes

Date: 17 November 2026

Approved by QA